
NOTES:

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- COATING (APPLY ACROSS COATING APERTURE)
  OPTICAL DENSITY = 0.6± 10% FROM 250 700
  \$1: PROPRITARY N.D.
  \$2: NONE
- 4. FINE GRIND SURFACE
- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION P-V @ 632.8nm:  $\leq \lambda/4$
- 7. ROHS COMPLIANT



FOR INFORMATION ONLY:



UV-VIS NEUTRAL DENSITY FILTER

## SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

SHAPE	S1 PLANO	S2 PLANO				Edmund Optics <sup>®</sup>
SURFACE QUALITY	40-20	40-20				OD 0.6, Ø50mm, UV-VIS ND FILTER
CLEAR APERTURE	Ø40	Ø40	THIRD ANGLE . PROJECTION	$\bigcirc \bigcirc$	TITLE	
COATING APERTURE	Ø40	Ø40		1		
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	62664 SHEE 1 OF