## NOTES:

- SUBSTRATE: Fused Silica
- 2. SURFACE \$1 TO BE PARALLEL TO SURFACE \$2 TO WITHIN <3 ARCMINS
- 3. COATING (APPLY ACROSS COATING APERTURE)


S1: R(ABS) >99.8% @ 355nm DAMAGE THRESHOLD, PULSED: 15 J/cm² @ 1030nm, 20ns, 20Hz


S2: NONE

4. FINE GROUND SURFACE

- 5. CLEAR APERTURE AND COATING APERTURE ARE CENTERED ON SURFACE
- 6. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY **ACROSS CLEAR APERTURE**

ARROW ON EDGE WITH LASER ETCH, PENCIL, OR PERMANENT INK POINTS **TOWARDS SURFACE S1** 





## PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2	
SHAPE	PLANO	PLANO	
SURFACE QUALITY	10-5	COMMERCIAL POLISH	
SURFACE FLATNESS	0.10 WAVE	N/A	
MIN CLEAR APERTURE	21.59 x 21.59	N/A	
MIN COATNG APERTURE	21.59 x 21.59	N/A	
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	

			<b>Edmund Optic</b>	S®
THIRD ANGLE PROJECTION	$\bigoplus \Box$	TITLE	25.4 x 25.4mm 355nm 0-45°, Nd:YAG Laser Line Mirror	
ALL DIMS IN	mm	DWG NO	39624	SHEET 1 OF 1